












尿中のカフェイン代謝物  
Caffeine metabolites in human urine

F041213B

A) UV at 255nm



B) UV at 295nm



Courtesy of Takashi FURUTA, Ph.D., Tokyo University of Pharmacy and Life science

Column : YMC-Pack Polyamine II  
250 × 4.6 mm I.D.  
Eluent : acetonitrile/0.75% formic acid (4/1)  
Flow rate : 1.0 mL/min  
Temperature : ambient  
Detection : A) UV at 255 nm, B) UV at 295nm  
Injection : Standard 10 μL  
Urine sample 10 μL